

Typical Specifications

UV-1900i

UV-VIS Spectrophotometer

The UV-1900i is a double-beam UV-Vis spectrophotometer using Shimadzu's original LO-RAY-LIGH™ diffraction grating technology. In addition to its high optical performance, the UV-1900i features high resolution, low stray light, high reproducibility, and an ultra-fast scan function. It also has an easy-to-use interface on a color touch-screen display. The UV-1900i is designed to meet the needs of both high performance and usability.

Hardware Specifications

ltem	Specification
Wavelength range	190 to 1,100 nm
Spectral bandwidth	1 nm (190 to 1,100 nm)
Wavelength setting	0.1 nm increments
	(1 nm increments when setting scanning range)
Wavelength accuracy	± 0.05 nm at D2 peak 656.1 nm,
	± 0.3 nm for entire range
Wavelength repeatability	± 0.025 nm
Wavelength slew rate	Approx. 29,000 nm/min
Wavelength scanning speed	3,000 to 2 nm/min
	29,000 nm/min when survey scanning
Lamp interchange wavelength	Automatic interchange linked to wavelength.
	The interchange wavelength can be set freely in
	the range of 295 to 364 nm (0.1 nm increments).
Stray light	Less than 0.004% at 220 nm (Nal)
	Less than 0.004% at 340 nm (NaNO ₂)
	Less than 0.15% at 198 nm (KCI)
Photometric system	Double beam optics
Photometric range	Absorbance: -4 to 4 Abs
	Transmittance: 0% to 400%
Photometric accuracy	± 0.0015 Abs at 0.5 Abs
	± 0.002 Abs at 1.0 Abs
	± 0.004 Abs at 2.0 Abs
	(measured using NIST930D/NIST1930 or equivalent.)
Photometric repeatability	Less than ± 0.00002 Abs at 0.5 Abs
	Less than ± 0.00003 Abs at 1 Abs
	Less than ± 0.00007 Abs at 2 Abs

The specifications shown here represent the average performance of the UV-1900i. These specifications are typical values, not guaranteed values. The guaranteed specifications are listed in a separate publication.



Navigate Your Way

ltem	Specification
Baseline stability	Less than 0.0002 Abs/Hr (700 nm,
	one hour after light source turned ON)
Baseline flatness	Less than ± 0.0003 Abs (1,100 to 190 nm,
	one hour after light source turned ON)
Noise level	Less than 0.00001 Abs (700 nm)
Light source	20-W halogen lamp and deuterium lamp
	Built-in light source auto position adjustment
Monochromator	LO-RAY-LIGH grade blazed holographic
	grating in Czerny-Turner mounting
Detector	Silicon photodiode
Sample compartment	Internal dimensions: W110 × D250 × H115 mm
	Distance between light beams: 100 mm
Power requirements	AC100,120,220,230,240 V,
	50/60 Hz, 140 VA
Environmental	Temperature: 15°C to 35°C
requirements	Humidity: 35% to 80%
	(without condensation; 70% max. at 30°C or higher)
Dimensions	W450 × D501 × H244 mm
Weight	16.6 kg
Output device	USB memory (optional)
	Extended memory (optional)
	Data files saved in text format or UVPC format*
	*Files in UVPC format can be read with the
	UVProbe file viewer, which is a function of
	LabSolutions™ UV-Vis, or with UVProbe software
PC compatibility	LabSolutions UV-Vis software (standard)
	External control possible via USB.
Display	24-bit color touch screen
	Touch pen (standard included)
	Touch panel protective sheet (optional)
Supported languages	Japanese, English, Chinese, Spanish (Mexico),
	Portuguese (Brazil), German, French, Russian.

Software Specifications

Measurement mode	Specification	Measurement mod
Spectrum mode	1. Measurement modes: ABS, T%, E	Biomethod mode
	2. Number of repeat scans: 1 to 99	
	3. Recording system: Selection between	
	single spectrum and data overlay	
	4. Data storage and recall	
	5. Data processing:	
	Peak/valley detection, arithmetic operations,	
	differentiation, smoothing, area calculation,	Maintenance
	point picking, data reading at	iviairiteriarice
	cursor-specified point	
Photometric mode	Single-wavelength measurement	
	1. Photometric modes: T% or Abs	
	2. Quantitation using K-factor method	
	3. Data table storage and recall functions	
	Multiple-wavelength measurement	
	4. Photometric modes: T% or ABS	
	5. Measurements at up to eight designated	
	wavelengths (set in 0.1 mm increments)	
	6. Data calculation at up to four	
	wavelengths (difference or ratio between	
	two wavelengths, calculation between	
	three wavelengths, etc.) is possible.	
Quantitation mode	1. Measurement methods:	
	1-wavelength, 2-wavelength, 3-wavelength,	
	and 1st to 4th derivative methods	
	2. Quantitation methods:	
	Automatic concentration calculation using K-factor	
	Automatic concentration calculation using	Shared functions
	single-point calibration curve	
	Multi-point calibration curve method	
	(1st to 3rd order regression curves)	
	3. Measurement parameters:	
	Number of standards (2 to 10)	
	Number of repeat measurements	
	(1 to 10 times) to obtain a mean value	
	for quantitation.	
Kinetics mode	Measures absorbance changes as a function of	
	time and calculates the enzymatic activity values.	
	2. Measurement time: 1 to 9,999 sec/min	
	Measurement methods: 1-wavelength,	
Timo scan mode	2-wavelength, multi-cell, and rate measurements 1. Measures changes in measured values as a function of time.	
Time scan mode	3	
	2. Measurement mode: ABS, T%, E	
	3. Measurement time: 1 to 9,999 sec/min	
	Data processing functions (same as spectrum mode)	

Measurement mode	Specification
Biomethod mode	DNA/Protein Quantitation 1. Calculation of DNA/protein concentration and absorbance ratio DNA concentration = K1 × A1 - K2 × A2 Protein concentration = K3 × A2 - K4 × A1
	Factors and measurement wavelengths can be set freely. Background correction is possible. Quantitation of proteins Quantitation methods: Lowry method, BCA method, Biuret
	method, CBB method (Bradford method), UV method
Maintenance	1. Baseline correction 2. Lamp usage time display and reset. 3. Security settings Functions can be restricted according to the user. 4. Instrument validation functions: 1) Compatible with 9 JIS items Wavelength accuracy, wavelength repeatability, resolution, stray light, photometric accuracy, photometric repeatability, baseline flatness, baseline stability, noise level. 2) Semi-automatic validation Validation inspections conducted interactively while inserting and removing inspection jigs. 3) Fully automatic validation Automatic validation Automatic validation inspections from measurement to evaluation and printout. 4) Setting inspection parameters and pass/fail criteria Authority to make changes can be protected by password access. 5) Detailed printout of results. 6) Bulk printout of results. 7) Equipped with method in accordance with
Shared functions	Pharmacopeia (JP, USP, EP). 1. Automatic setting of measurement mode after instrument initialization. It's possible to specify standby and parameter files in the parameter setting window for each measurement mode. 2. Selection of displayed number of decimal places Absorbance: 3 or 4 decimal places Transmittance: 1 or 2 decimal places 3. Number of files that can be saved (built-in memory) Measurement parameters: 100 files max. Tabular data: 15 files max. Curve data: 16 files max. Validation condition: 10 files max. Validation result: 3 files max. 4. Number of savable files (Extended memory and options) Curve data files: 999 files max. 5. Keyboard and Bar code Reader Support. File names can be entered using the keyboard or bar code reader. 6. Wake-up function Equipment can be started at a fixed time every day. 7. Network Connections Multiple UV -1900i can be connected to a single printer over the network. 8. Setting of integration time (for fixed-wavelength measurement) PC control Spectrophotometer can be controlled by an external PC. This function is also used when performing operation with the standard LabSolutions UV-Vis software provided.

LO-RAY-LIGH and LabSolutions are trademarks of Shimadzu Corporation.



Shimadzu Corporation www.shimadzu.com/an/

For Research Use Only. Not for use in diagnostic procedures.
This publication may contain references to products that are not available in your country. Please contact us to check the availability of these products in your country.
Company names, products/service names and logos used in this publication are trademarks and trade names of Shimadzu Corporation, its subsidiaries or its affiliates, whether or not they are used with trademark symbol "TM" or "®".
Third-party trademarks and trade names may be used in this publication to refer to either the entities or their products/services, whether or not they are used with trademark symbol "TM" or "®".
Shimadzu disclaims any proprietary interest in trademarks and trade names other than its own.

Shimadzu does not assume any responsibility or liability for any damage, whether direct or indirect, relating to the use of this publication.

*A USB cable is required.